



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Abbott recalls Alinity S Software for Error during Immunoassay Wash Cycle

Recall Date	Product Description	Recalling Firm	Recall Reason
02/03/2021	Alinity s System software version 2.8.0, REF LN 06P16-01, IVD, OEI: (01)0038074-0138479 The Alinity s System is intended for In Vitro diagnostic use only. The Alinity s System is a highthroughput, fully automated immuneassay analyzer designed to determine the presence of specific antigens and antibodies by using chemiluminescent immunoassay technology.	Abbott Laboratories Irving, Texas	Software error associated with the immunoassay analyzer wash cycle which is using 1 mL of wash buffer instead of the intended 3 mL of wash buffer to wash the exterior of the probe.

Recall Class	Product Identification	Distribution	Affected Dates
II	Serial numbers AS1001 through AS1387 UDIs: (01)00380740138479(21)AS1002 (01)00380740138479(21)AS1005 (01)00380740138479(21)AS1009 (01)00380740138479(21)AS1012 (01)00380740138479(21)AS1013 (01)00380740138479(21)AS1016	4 Units affecting 3 Facilities	February 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

