



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Accuray Incorporated Recalls The Upper Roller Bracket For Lack of Enclosure Space**

Recall Date	Product Description	Recalling Firm	Recall Reason
9/6/2023	<b>Upper roller bracket</b> (Part number 1088370) for the Gen2 gantry enclosure for the Radixact Treatment Delivery System	<b>ACCURAY INCORPORATED</b> Madison, Wisconsin	It is possible the upper roller bracket in the Radixact System gantry enclosure may not provide enough clearance for service personnel to use a jib crane for moving components while servicing the system, resulting in possible injury to the service personnel.

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>Upper Roller Bracket</b> UDI-DI number 00811376030009. Serial numbers 4010505, 4010506, 4010507, 4010508, 4010509, 4010510, 4010511, 4010513, 4010515, 4010516, 4010517, ...	82 Units Nationwide including California	August 3, 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

