



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Activbody recalls Activ 5 Tiny Gym and ActivForce2 for Inaccurate Force Measurement and Registration Error**

Recall Date	Product Description	Recalling Firm	Recall Reason
11/30/2021	<b>activ5 TINY GYM IN YOUR POCKET</b>	<b>Activbody, Inc.</b> San Diego, California	Manufacturing issue results in inaccurate force measurement which may lead to incorrect treatment planning. Firm also failed to register and list its medical devices with FDA.
11/30/2021	<b>activforce 2</b> Activ5-M R 201-200563	<b>Activbody, Inc.</b> San Diego, California	Manufacturing issue and registration error

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>activ5 TINY GYM IN YOUR POCKET</b> Activforce Kit Number: AB0000-AC5540 Serial Numbers: 5P1011011708AB0000 -5P1011011708AB1712	58 Units in California	November 2021 and prior
II	<b>activforce 2</b> Serial Numbers: 5M1011012046AK7655 - 5M1011012102AL9371	15 Units in California	November 2021 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

