



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Ansell Healthcare Products Recalls Gammex Non-Latex Pi Green Surgical Gloves For Unsealed Sterile Pouches

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|-------------|---|---|---|
| 10/18/2023 | GAMMEX NON-LATEX PI GREEN Surgical Gloves REF: 20685255 through 20685290 | ANSELL HEALTHCARE PRODUCTS LLC Reno, Nevada | Some surgical glove sterile pouches were not completely sealed. |
| 10/18/2023 | Novaplus Gammex Non-Latex Pi Green Surgical Gloves REF: V20685255 through V20685290 | ANSELL HEALTHCARE PRODUCTS LLC Reno, Nevada | Some surgical glove sterile pouches were not completely sealed. |

| Recall Class | Product Identification | Distribution | Affected Dates |
|--------------|---|---|-------------------------|
| II | GAMMEX NON-LATEX PI GREEN Surgical Gloves Lot: 2203434204, REF/UDI-DI: 20685255/00884792385206, 20685260/00884792385213, 20685265/00884792385220, 20685270/00884792385237, 20685275/00884792385244, 20685280 00884792385251, 20685285 00884792385268, 20685290 00884792385275. | 1511 Units Nationwide including California | 3 August 2023 and prior |
| II | NOVAPLUS GAMMEX NON-LATEX PI GREEN Surgical Gloves Lot: 2203434604, REF/UDI-DI: V20685255/10014566400553, V20685260/10014566400607, V20685265/10014566400652, | 378 Units Nationwide including California | 3 August 2023 and prior |

| | | | |
|--|---|--|--|
| | V20685270/10014566400706, V20685275/10014566400751, V20685280/10014566400805, V20685285/10014566400850, V20685290/10014566400904. | | |
|--|---|--|--|

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

