



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Becton Dickinson Infusion Therapy Systems Recalls BD Nexiva for Breach

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|-------------|---|--|--|
| 9/8/2021 | BD Nexiva, 20 GA 1.00 IN (1.1 X 25 mm), 3660 ml/hr. 61ml/min; Closed IV Catheter System- Over-the-needle, intravascular catheters. | Becton Dickinson Infusion Therapy Systems Inc. Sandy, Utah | There is a breach in the product packaging that renders the product non-sterile. |

| Recall Class | Product Identification | Distribution | Affected Dates |
|--------------|---|--|-----------------------------|
| II | Catalog number/Model number: 383536; Batch numbers/UDI numbers: Lot: 1193055/(17)24 0630(10)1193055(30)80/(01)50 382903835367, 2024-06-30; Lot: 1166273: (17)240630(10) 1166273(30)80/(01)503829038 35367, 2024-06-30 Lot: 1166 273: (17)240630(10)1166273 (30)80/(01)50382903835367, 2024-06-30 Lot: 1188953: (17) 240630(10)118895 3(30)80/(01))50382903835367, 2024-06-30 | 155,840 devices, 80 units in each box Nationwide including California | September 2021 and prior |

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

