



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Cook Vandergrift Recalls Cook Lead Clipper For Compromised Sterilization

Recall Date	Product Description	Recalling Firm	Recall Reason
9/13/2023	Cook Lead Clipper An auxiliary tool indicated for use in patients requiring percutaneous retrieval of cardiac leads. Order Number (GPN): G20003 Reference Part Number (RPN): LR-CLP001	COOK VANDERGRIFT, INC. Vandergrift, Pennsylvania	Sterility of device may be compromised due to breach of the chevron seal of the packaging

Recall Class	Product Identification	Distribution	Affected Dates
II	Cook Lead Clipper Device Identifier (DI): (01)00827002200036(17) Lot Numbers: N172707 N172737 N172738 N172875 N172887 N172912 N173153 N173178 N173180 N173434 N173519 N173623 N173653 N173688 N173690 N173691 N173879 N173943 N173959 N173976 N173977 ...	15951 Units Nationwide including California	24 July 2023, and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

