



Food and Drug Branch – Device Recalls

California Device Recall Information Sheet

Covidien Lp Recalls Absorbatack Absorbable Fixation Device Due To Low Seal Strength Of The Side Seal Of The Foil Pouch Of The Device

Recall Date	Product Description	Recalling Firm	Recall Reason
2/14/2024	Absorbatack Absorbable Fixation Device 30 Violet Absorbable Tacks 5 mm- -intended for fixation of prosthetic material to soft tissue in minimally invasive and open surgical hernia repair procedures Model ABSTACK30	Covidien LP Mansfield, Massachusetts	This device is being recalled due to the packaging of the device causing low seal strength in the side seal of the foil pouch, potentially compromising the product sterility.

Recall Class	Product Identification	Distribution	Affected Dates
II	Absorbatack Absorbable Fixation Device GTIN 20884523006527 Lot number: N3J1860Y	18 Units in California California	February, 2024 and prior

For additional information, please visit the [FDA website](#)

