



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Draeger Medical Recalls Seattle PAP Plus And Breathing Circuit/Anesthesia Kits For Risk Of Loose Or Detached Components That Can Restrict Breathing Support

Recall Date	Product Description	Recalling Firm	Recall Reason
5/24/2023	Seattle Pap Plus Ventilator, non-continuous (respirator) MP02260	DRAEGER MEDICAL Telford, Pennsylvania	Glued connections of the breathing circuit can become loose before or during ventilation resulting in partial or complete detachment of pieces.

Recall Class	Product Identification	Distribution	Affected Dates
I	Product 1 Product 1 Identification Details UDI etc	10705 Units in California	March 2023 - March 2020

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

