



State of California—Health and Human Services Agency  
**California Department of Public Health**



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Draeger Medical, Inc. Oxylog 3000 Plus for Device Malfunctions**

Recall Date	Product Description	Recalling Firm	Recall Reason
7/19/2023	<b>Oxylog 3000 Plus</b> Emergency and transport ventilator	<b>Draeger Medical, Inc.</b> Telford, PA	Draeger has become aware of instances where the device stopped working when in use when transitioning from battery to AC operation.

Recall Class	Product Identification	Distribution	Affected Dates
I	<b>Oxylog 3000 Plus</b> Catalog No. 5704811 & 5704813; UDI-DI: 04048675398295; All Serial No.	17 Units in California	June 2023 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch  
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435  
 (916) 650-6500 • (916) 650-6650 FAX  
 Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

