



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Draeger Medical Systems, Inc. Softbed Resuscitaire and or Birthing Room Warmer with Scale for Inaccurate Weight Values

Recall Date	Product Description	Recalling Firm	Recall Reason
6/28/2023	Draeger Softbed Resuscitaire and/or Birthing Room Warmer with Scale [Model(s): RW82VHA-1, RW82VHA-1C, RW82VHA-1D, RW82-1, RW82-1C, RW82-1D, WBR82-1 and WBR82-1C]	Draeger Medical Systems, Inc. Telford, Pennsylvania	The Resuscitaire Infant Radiant Warmer with the optional scale could display inaccurate weight values.

Recall Class	Product Identification	Distribution	Affected Dates
II	Draeger Softbed Resuscitaire and/or Birthing Room Warmer with Scale All units shipped from January 20, 2020 through March 1, 2023; MU20503 (Resuscitaire) UDI - 04049098000215 MU20512 (Birthing Room Warmer - BRW) UDI - 04049098058001	11 Units in California	May 2023 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

