



**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**GE Healthcare recalls Backup Batteries for Carescape Ventilators for Insufficient Power**

Recall Date	Product Description	Recalling Firm	Recall Reason
4/18/2022	<b>Backup Batteries for Carescape R860 Ventilators</b> manufactured on or after April 1, 2019	<b>GE Healthcare, LLC</b> Waukesha, Wisconsin	Insufficient battery backup power resulting in premature shutdown of the ventilator when not connected to AC mains power supply.
4/18/2022	<b>Backup Batteries for Carescape R860, Engstrom Carestation and Engstrom Pro Ventilators</b> distributed on or after April 1, 2019	<b>GE Healthcare, LLC</b> Waukesha, Wisconsin	Insufficient battery backup power

Recall Class	Product Identification	Distribution	Affected Dates
I	<b>Backup Batteries for Carescape R860 Ventilators</b> Model Number: 1506-8600-000; UDI/DI Number:00840682102346; Serial Numbers: CBRZ06959 - WBRA00574	Nationwide including California	April 2019 to April 2022
I	<b>Backup Batteries for Carescape R860, Engstrom Carestation and Engstrom Pro Ventilators</b> Field Replacement Unit kits PN: 1009-5682-000-S (BTRY SEALED LEAD ACID RECHARGEABLE 12V) and PN: 5856787-S (BTRY SEALED LEAD ACID RECHARGEABLE 12V PAIR).	Nationwide including California	April 2019 to April 2022

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](https://www.fda.gov)

