



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

GE Healthcare Recalls Aisys and Avance Anesthesia Systems

Recall Date	Product Description	Recalling Firm	Recall Reason
6/01/2022	Aisys CS2 Anesthesia System Part number 1011-9050-000	GE Healthcare, LLC Waukesha, Wisconsin	There is a potential reversal of the O2 and air cylinder pressure transducer connections.
6/01/2022	Avance CS2 and Avance CS2 Pro Anesthesia System Part number 1009-9050-000	GE Healthcare, LLC Waukesha, Wisconsin	There is a potential reversal of connections.

Recall Class	Product Identification	Distribution	Affected Dates
I	Aisys CS2 GTIN 00840682102292 Serial numbers starting with APWA: 00101 through 02539 APWB: 00101 through 00952 APWZ: 03034 through 03422	5,372 Devices Nationwide including California	June 2022 and prior
I	Avance CS2 and Avance CS2 Pro GTIN 00840682102322 APKA: 00101 through 02029 APKB: 00101 through 00517 APKZ 00247 through 02569		June 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

