



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Intera Oncology Recalls Intera 3000 Hepatic Artery Infusion Pump

Recall Date	Product Description	Recalling Firm	Recall Reason
8/31/2022	INTERA 3000 Hepatic Artery Infusion Pump 30 mL, Sterile, Rx Only. implantable infusion pump	Intera Oncology, Inc. Wellesley, Massachusetts	Higher than expected flow rate.

Recall Class	Product Identification	Distribution	Affected Dates
I	UDI-DI: 00850014110147; Catalogue No. AP03000H; Serial No. 16145 – 16485	50 units in multiple states, including California	July 2022 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

