



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Karl Storz Endoscopy Recalls CMOS Video Cysto-Urethroscope for Image Loss**

Recall Date	Product Description	Recalling Firm	Recall Reason
4/01/2022	<b>CMOS Video Cysto-Urethroscope</b> STORZ KARL STORZ- ENDOSKOPE REF 11272VUE CMOS Video Cysto- Urethroscope REF 11272VUEK CMOS Video Cysto-Urethroscope C-VIEW KIT REF 11272VUE-R CMOS Video Cysto-Urethroscope REF 11272VUEK-R CMOS Video Cysto-Urethroscope C- VIEW KIT	<b>Karl Storz            Endoscopy</b> El Segundo, California	The endoscope may experience a loss of image when activating monopolar electrocautery, due a manufacturing assembly process issue.

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>Video Cysto-Urethroscope</b>	7 units in California	April 2022 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](https://www.fda.gov)

CDPH Food and Drug Branch  
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435  
 (916) 650-6500 • (916) 650-6650 FAX  
 Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

