



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Medline Industries Recalls Hudson Rci Addipak Unit Dose Vials For Sterility Failure

Recall Date	Product Description	Recalling Firm	Recall Reason
9/20/2023	Medline Industries HUDSON RCI Addipak, UNIT DOSE VIALS, 5mL Sterile 0.9% NaCl Solution for Inhalation USP. REF RHUD59U	Medline Industries Northfield, Illinois	Product dispositioned for scrap for sterility failure investigation that was inadvertently shipped to customers.

Recall Class	Product Identification	Distribution	Affected Dates
I	Medline Industries HUDSON RCI Addipak, UNIT DOSE VIALS, UDI/DI 140267047147818, Lot code 3B085	18,000 Units Nationwide	August 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

