



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Medtronic Recalls Sensight Connector Plug For Mislabeled Use-By-Date**

Recall Date	Product Description	Recalling Firm	Recall Reason
10/25/2023	<b>Sensight Connector Plug</b> Model Number B31061	<b>MEDTRONIC NEUROMODULATION</b> Minneapolis, Minnesota	Medtronic is requesting return of a small number of units of the SenSight Connector Plug, Model B31061, due to a discrepancy in the Use By Date (UBD) printed on the label.

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>Sensight Connector Plug</b> GTIN 00763000426576, Lot Number 082U30721	65 Units Nationwide including California	Manufactured 1/20/2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

