



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Nurse Assist Llc Recalls Iv And Injectable Products For Sterility

Recall Date	Product Description	Recalling Firm	Recall Reason
12/27/2023	Mac Medical 0.9% Sodium Chloride 10ML FILL SYRINGE/9210; NURSE ASSIST: 3ML IV FLUSH SYRINGE/1203-BP, 5ML IV FLUSH SYRINGE/1205-BP, 10ML IV FLUSH SYRINGE/1210-BP; SOL: 0.9% SODIUM CHLORIDE IV FLUSH	NURSE ASSIST, LLC Haltom City, Texas	Sodium chloride irrigation USP, and sterile water for irrigation USP, that may be packaged inside kits, have the potential for a lack of sterility assurance, which could result in a non-sterile product, use of which could cause infection.
12/27/2023	Sterile Water USP STERILE WATER SYRINGE/1030A, SYRINGE PREFILLED 10 ML/6496722, 30CC PREFILLED SYRINGE/6496723, USP STERILE WATER SYRINGE/1030, 10CC PREFILLED SYRINGES/51638, 30CC PREFILLED SYRINGES/52333, 10CC PREFILLED SYRINGES/1010	NURSE ASSIST, LLC Haltom City, Texas	lack of sterility assurance
12/27/2023	Sterile Water/Saline CUP STERILE WATER, FOIL LID 120 ML/T167005; CUP STERILE SALINE FOIL LID 120 ML/T167007	NURSE ASSIST, LLC Haltom City, Texas	lack of sterility assurance
12/27/2023	Sterile Water/Saline CARDINAL: 100ML STERILE	NURSE ASSIST, LLC Haltom City, Texas	lack of sterility assurance

SALINE BOTTLE/1020, 100ML
STERILE SALINE
BOTTLE/1022, 100ML
STERILE WATER
BOTTLE/1024, 100ML
STERILE WATER
BOTTLE/10000, 100ML
STERILE WATER
BOTTLE/PT00103419, 100ML
STERILE SALINE
BOTTLE/PT00103427;
COVIDIEN: STERILE WOUND
SOLUTION/90SAL, STERILE
WOUND SOLUTION/210SAL;
HAYLARD OWENS MINOR:
250ML STERILE WATER
BOTTLE/49148, 250ML
STERILE SALINE
BOTTLE/49152, 1000ML
STERILE SALINE
BOTTLE/49154, 100ML
STERILE SALINE
BOTTLE/49155, 1000ML
STERILE WATER
BOTTLE/49150, 100ML
STERILE WATER
BOTTLE/49151, 500ML
STERILE WATER
BOTTLE/49149, 500ML
STERILE SALINE
BOTTLE/49153; IDEXX:
BOTTLE STERILE WATER 10
PACK/98-09444-01;
MCKESSON: 100ML STERILE
SALINE BOTTLE/37-6240,
500ML STERILE WATER
BOTTLE/37-6290, 250ML
STERILE SALINE BOTTLE/37-
6270, 500ML STERILE SALINE
BOTTLE/37-6280, 120ML
STERILE WATER CUP/37-
6210, 1000ML STERILE
SALINE BOTTLE/37-6281,
1000ML STERILE WATER
BOTTLE/37-6291, 120ML
STERILE SALINE CUP/37-
6220, 100ML STERILE
WATER BOTTLE/37-6250,
250ML STERILE WATER
BOTTLE/37-6260, 210ML
SALINE WOUND FLUSH/37-
6507, 88.7ML SALINE WOUND
FLUSH STERILE/37-6503;
MEDICHOICE OWENS

MINOR: CAN STERILE
SALINE SPRY MEDI 7.1
OZ/SWF071; MEDLINE:
110ML STERILE SALINE
CUP/DYND40540, 100ML
STERILE SALINE
BOTTLE/RDI30296, 250ML
STERILE WATER
BOTTLE/PCS1550, 250ML
STERILE SALINE
BOTTLE/PCS1650, 1000ML
STERILE SALINE
BOTTLE/DYND1000MLS,
100ML STERILE SALINE
BOTTLE/52011, 500ML
STERILE WATER
BOTTLE/DYND500MLW,
500ML STERILE SALINE
BOTTLE/DYND500MLS,
100ML STERILE WATER
BOTTLE/RDI30295, CAN
SALINE STERILE SPRAY
CURAD/CURSALINE7, 110ML
STERILE WATER
CUP/DYND40570, 100ML
STERILE WATER
BOTTLE/54380; NURSE
ASSIST: 100ML STERILE
SALINE BOTTLE/6492775,
100ML STERILE WATER
BOTTLE/6492776;
STERICARE: 90ML USP
SALINE WOULD FLUSH/6503,
100ML STERILE SALINE
BOTTLE/6240, 100ML
STERILE WATER
BOTTLE/6250, 250ML
STERILE SALINE
BOTTLE/6270, 500ML
STERILE WATER
BOTTLE/6290, 250ML
STERILE WATER
BOTTLE/6260, 500ML
STERILE SALINE
BOTTLE/6280, 1000ML
STERILE SALINE
BOTTLE/6281, 1000ML
STERILE WATER
BOTTLE/6291, 210ML SALINE
WOUND FLUSH/6507, 120ML
STERILE SALINE FOIL LID
CUP/6220, 120ML STERILE
WATER FIOL LID CUP/6210,
500ML STERILE WATER

	BOTTLE/5290, 500ML STERILE SALINE BOTTLE/5280; VYAIR: CUP STERILE SALINE FOIL LID 120 ML/4873C		
12/27/2023	0.9% Sodium Chloride 10ML IV FLUSH STERILE FIELD/EMZ10091240	NURSE ASSIST, LLC Haltom City, Texas	lack of sterility assurance

Recall Class	Product Identification	Distribution	Affected Dates
I	Mac Medical Product Code/UDI-DI/Lot: 9210/20850020865042/211225 20-22032938; 1203-BP /B1501203BP4/22022797; 1205-BP/B1501205BP4 /22022789, 23024788; 1210- BP/B1501210BP4/21112359- 23096407; PFF001/ 20810062438895/23014656, 23014671; SS003/ 20885632389778/23035163- 23086334; SS005/ 20885632389761/23035151- 23086323; SS010/ 20885632389754/23024926- 23086264; SS010SFR/not distributed.	61,740 Cases Units Nationwide including California	December 2023 and prior
I	Sterile Water Product Code/UDI-DI/Lot: 1030A/B1501030A4/2001155- 23034943; 1030/B15010304 /22032995, 23014586; 51638 /B150516384/22033023, 23014632; 52333/B150523334 /22033017; 1010/B15010104 /22063386-23014593	15,800 Cases Units Nationwide including California	December 2023 and prior

I	<p>Sterile Water/Saline Product Code/Lot: T167005/22022859, 22083878, 23055524; T167007/21122503, 22022860, 22083862, 22114365, 23045255, 23045267</p>	432 Units Nationwide including California	December 2023 and prior
I	<p>Sterile Water/Saline - Cardinal Product Code/UDI-DI/Lot: 1020/50192253044490/211226 19-23076080; 1022/ 50192253044506/22012651- 23076088; 1024/ 50192253044513/21122618- 23076098; 10000/ 50192253044520/22073713- 23076079; PT00103419/ 50192253044513/23076122; PT00103427/50192253044506/ 23076140, 23086258; 90SAL/ 20884521140940/22083848; 210SAL/10884521139756/2202 2771-22114334; 49148/ 20885632395878/23076067; 49152/20885632395915/23076 041; 49154/20885632395939/ 23076066; 49155/ 20885632395946/23076057; 49150/20885632395892/23076 074; 49151/20885632395908/ 23076040; 49149/ 20885632395885/23086163 49153/20885632395922/23086 156; 98-09444-01/21112440- 23096487; 37-6240/ 612479168565/21112357- 23086378; 37-6290/ 612479180314/21102308- 23086253; 37-6270/ 612479168596/21102302- 23096424; 37-6280/ 612479180307/21102305- 23086261; 37-6210/ 612479168541/21102252- 23096453; 37-6281/ 612479228399/21112452- 23086337; 37-6220/ 612479168558/21122549- 23096425; 37-6250/ 612479168572/22012640- 23086289; 37-6260/</p>	31850 Units Nationwide including California	December 2023 and prior

	<p>612479168589/22022737-23076017; 37-6503/ 612479228429/22083835; SWF071/50885632127043/22022738-23076008; DYND40540/40884389334233/21112358-23055699; RDI30296/ 40884389338248/21112391-23055681; PCS1550/ 40080196427056/21112394-23055719; PCS1650/ 40884389338262/21112398-23065745; DYND1000MLS/ 208882776699466/21112415/21112411; 52011/21112453, 22033018-22093967; DYND500MLW/40884389118390/21112481-23034983; DYND500MLS/40884389118406/21112387-23034984; RDI30295/40884389338231/22012724-23045397; CURSALINE7/40080196297970/22032903-22043097; DYND40570/40884389334226/22053264-23055676; 6492775/B15064927754G/21112412-22083859; 6240/ 10850421008812/21112395-23076042; 6250/ 10850421008768/21112397-23096475; 6290/ 10850421008782/21102214-23096482; 6260/ 10850421008775/21092009-23096440; 6280/ 10850421008836/21122490-23096469; 5290/850421008914/23065731-23065892; 5280/ 850421008907/23065735-23065919; 4873C/21112462</p>		
I	<p>0.9% Sodium Chloride Product Code/UDI-DI/Lot: EMZ10091240/20363807109166/21112364-22104234</p>	3360 Units Nationwide including California	December 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#) for [Mac Medical](#), [USP Sterile Water Syringe](#), [Cup Sterile Water](#), [CARDINAL 100mL Sterile Saline Bottle](#), [10mL IV Flush Sterile Field](#)

