



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Olympus Corporation Of The Americas Recalls The Air/Water Valve Provided/Used With The Endoscopes For the loss of one-way valve functionality.**

Recall Date	Product Description	Recalling Firm	Recall Reason
8/30/2023	<b>Olympus Corporation Of The Americas</b> The Air/Water Valve provided/used with the following Endoscopes: ULTRASONIC GASTRO-FIBERSCOPE, ULTRASONIC GASTROVIDEOSCOPE, ULTRASONIC COLONOV-IDEOSCOPE: GF-UC140P-AL5, GF-UCT140-AL5, GF-UE160-AL5, GF-UCT180, GF-UM20, GF-UM130, GF-UMQ130, GF-UM160, GF-UC160P-OL5,GF-UCT160-OL5	<b>Olympus Corporation Of The Americas</b> Center Valley, Pennsylvania	The air/water valve MAJ-1444 used with OER-Pro and OER-Elite may become damaged and result in loss of the oneway valve functionality.

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>Olympus Corporation Of The Americas.</b> The Air/Water Valve UDI-DI: 04953170355929. All lot numbers	29590 Units Nationwide	July 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch  
 MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435  
 (916) 650-6500 ● (916) 650-6650 FAX  
 Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

