



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Acumed Llc. Recalls Ribloc U Plus 90 Instrument Primary Guide Assy, Low Profile Ref Rbl2320 For Potential Breakage During Use.

Recall Date	Product Description	Recalling Firm	Recall Reason
01/10/2024	<p>Ribloc U Plus 90 Instrument Primary Guide Assy, Low Profile Ref Rbl2320</p> <p>The RBL2320, Low Profile Primary Guide is an instrument used with a power system to compress the RibLoc plates onto the bone, determine the length of the screw to install, locate the position of the holes to drill, and provide proper orientation for the installation of the screws.</p>	<p>Acumed LLC Hillsboro, Oregon</p>	<p>Due to potential breakage during use.</p>

Recall Class	Product Identification	Distribution	Affected Dates
II	<p>Ribloc U Plus 90 Instrument Primary Guide Assy, Low Profile Ref Rbl2320</p> <p>Part Number: RBL2320 UDI Code / Batch-Lot Number 10806378118210-538351000000 / 538351 10806378118210-538352000000 / 538352 10806378118210-544036000000 / 544036 10806378118210-554505000000 / 554505 10806378118210-L1810012181031 / L1810012 10806378118210-L1902009190207 / L1902009 10806378118210-L1909017190918 / L1909017</p>	<p>23 Units in California</p>	<p>January 2024 and prior</p>

	10806378118210- L2001005200130 / L2001005 10806378118210- L2005002200504 / L2005002 10806378118210- L2010046000000 / L2010046 10806378118210- L2010048000000 / L2010048		
--	--	--	--

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [Ribloc U Plus 90 Instrument Primary Guide Assy](#)

