



CALIFORNIA DEVICE RECALL INFORMATION SHEET

Siemens Healthcare Diagnostics Recalls ADVIA Chemistry Urinary/Cerebrospinal Fluid Protein

Recall Date	Product Description	Recalling Firm	Recall Reason
08/30/2023	Siemens Healthcare Diagnostics. ADVIA Chemistry Urinary/Cerebrospinal Fluid Protein (UCFP), Material Number 11319151	Siemens Healthcare Diagnostics Tarrytown, New York	There is a potential for ADVIA Chemistry Urinary/Cerebrospinal Fluid Protein reagent carryover impacting Enzymatic Creatinine ₂ (ECRE ₂). Falsely depressed ECRE ₂ results may be observed.

Recall Class	Product Identification	Distribution	Affected Dates
II	Siemens Healthcare Diagnostics. ADVIA Chemistry Urinary/Cerebrospinal Fluid Protein (UCFP)UDI-DI: 00630414279176; Lot Numbers: All lots	3067 Units Nationwide including California	July 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435
 (916) 650-6500 • (916) 650-6650 FAX
 Internet Address: www.cdph.ca.gov

