



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Skytron recalls FS32 Monitor Bracket for Potential to Loosen and Detach from Baseplate

Recall Date	Product Description	Recalling Firm	Recall Reason
06/23/2021	Skytron Bracket Flatscreen Revision 2 & 3 The FS32 is a single monitor mounting bracket intended for use with a 32 inch (maximum) flat screen monitor. Model: FS32	Skytron, Div. The KMW Group, Inc Grand Rapids, Michigan	Failure of the FS32 Monitor Brackets from the mounting block could potentially loosen over time and eventually detach from the baseplate, result in the display interface dismounting from its attachment point, causing injury to patients or staff

Recall Class	Product Identification	Distribution	Affected Dates
II	Skytron Bracket Flatscreen 1811-314-00-1-0072 1811-314-00-1-0074 1811-314-00-1-0077 1811-314-00-1-0080	4 Units in California	December 2021 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

