



California Device Recall Information Sheet

Sleepnet Corporation Recalls Vented And Non Vented Masks For Updated Device Contraindications Associated With Containing Magnets

Recall Date	Product Description	Recalling Firm	Recall Reason
4/10/2024	Mojo Full Face With Headgear All Sizes	SLEEPNET CORPORATION Hampton, New Hampshire	Update to contraindications and warning language due to CPAP masks containing magnets.
4/10/2024	Mojo Non Vented Full Face Mask With Headgear All Sizes	SLEEPNET CORPORATION Hampton, New Hampshire	Update to contraindications and warning language due to CPAP masks containing magnets.
4/10/2024	Mojo 2 Full Face Vented Mask All Sizes	SLEEPNET CORPORATION Hampton, New Hampshire	Update to contraindications and warning language due to CPAP masks containing magnets.
4/10/2024	Mojo 2 Full Face Non Vented Mask All Sizes	SLEEPNET CORPORATION Hampton, New Hampshire	Update to contraindications and warning language due to CPAP masks containing magnets.
4/10/2024	Mojo 2 Full Face Aav Non Vented Mask All Sizes	SLEEPNET CORPORATION Hampton, New Hampshire	Update to contraindications and warning language due to CPAP masks

			containing magnets.
4/10/2024	Iq 2 Nasal Vented Mask	SLEEPNET CORPORATION Hampton, New Hampshire	Update to contraindications and warning language due to CPAP masks containing magnets.
4/10/2024	Phantom 2 Nasal Vented Mask	SLEEPNET CORPORATION Hampton, New Hampshire	Update to contraindications and warning language due to CPAP masks containing magnets.

Recall Class	Product Identification	Distribution	Affected Dates
I	Mojo Full Face With Headgear All UDI-DI; All Lot Numbers.	23042 Units Worldwide	March, 2024 and prior
I	Mojo Non Vented Full Face Mask With Headgear All UDI-DI; All Lot Numbers.	11874 Units Worldwide	March, 2024 and prior
I	Mojo 2 Full Face Vented Mask All UDI-DI; All Lot Numbers.	2787 Units Worldwide	March, 2024 and prior
I	Mojo 2 Full Face Non Vented Mask All UDI-DI; All Lot Numbers.	2107 Units Worldwide	March, 2024 and prior
I	Mojo 2 Full Face Aav Non Vented Mask All UDI-DI; All Lot Numbers.	1268 Units Worldwide	March, 2024 and prior
I	Iq 2 Nasal Vented Mask All UDI-DI; All Lot Numbers.	2477 Units Worldwide	March, 2024 and prior
I	Phantom 2 Nasal Vented Mask All UDI-DI; All Lot Numbers.	2417 Units Worldwide	March, 2024 and prior

For additional information, please visit the FDA website [Mojo with Headgear](#), [Mojo Non Vented with Headgear](#), [Mojo 2 Vented](#), [Mojo 2 Non Vented](#), [Mojo 2 AAV](#), [iQ 2](#), [Phanton 2](#)

CDPH Food and Drug Branch
MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435
(916) 650-6500 ● (916) 650-6650 FAX
Internet Address: www.cdph.ca.gov

