



## CALIFORNIA DEVICE RECALL INFORMATION SHEET

### Smith & Nephew Recalls Trigen Trochanteric Antegrade Nail Lime.

Recall Date	Product Description	Recalling Firm	Recall Reason
08/30/2023	<b>Smith &amp; Nephew, Inc.</b> TRIGEN TROCHANTERIC ANTEGRADE NAIL LIME 11.5MM X 36CM 130 DEGREE LEFT, REF 71647336, Femur nail	<b>Smith &amp; Nephew, Inc.</b> Memphis, Tennessee	A complaint was received indicating that a package contained a TRIGEN TROCHANTERIC ANTEGRADE NAIL LIME 11.5MM X 40CM 130 DEGREE LEFT implant instead of a TRIGEN TROCHANTERIC ANTEGRADE NAIL LIME 11.5MM X 36CM 130 DEGREE LEFT.
08/30/2023	<b>Smith &amp; Nephew, Inc.</b> TRIGEN TROCHANTERIC ANTEGRADE NAIL LIME 11.5MM X 40CM 130 DEGREE LEFT, REF 71647340, Femur nail	<b>Smith &amp; Nephew, Inc.</b> Memphis, Tennessee	A complaint was received indicating that a package contained a TRIGEN TROCHANTERIC ANTEGRADE NAIL LIME 11.5MM X 40CM 130 DEGREE LEFT implant instead of a TRIGEN TROCHANTERIC ANTEGRADE NAIL LIME 11.5MM X 36CM 130 DEGREE LEFT.

Recall Class	Product Identification	Distribution	Affected Dates
II	<b>Smith &amp; Nephew, Inc.</b> TRIGEN TROCHANTERIC ANTEGRADE NAIL, UDI/DI 03596010500922, Batch Number 21KSM0598	30 Units Nationwide including California	July 2023 and prior
II	<b>Smith &amp; Nephew, Inc.</b> TRIGEN TROCHANTERIC ANTEGRADE NAIL, UDI/DI 03596010500922, Batch Number 21KSM0606	32 Units Nationwide including California	July 2023 and prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch  
 MS 7602 • P.O. Box 997435 • Sacramento, CA 95899-7435  
 (916) 650-6500 • (916) 650-6650 FAX  
 Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

