



**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Stryker Corporation Recalls Cinchlock Ss Knotless Anchor Inserter For Distribution Of Expired Product**

| Recall Date | Product Description  | Recalling Firm                                     | Recall Reason               |
|-------------|--|--|-----------------------------|
| 10/4/2023   | <b>Cinchlock Ss Knotless Anchor Inserter</b><br>Part Number CAT02462 | <b>STRYKER CORPORATION</b><br>San Jose, California | Expired Product distributed |

| Recall Class | Product Identification   | Distribution                                | Affected Dates           |
|--------------|--|---|--------------------------|
| II           | <b>Cinchlock Ss Knotless Anchor Inserter</b><br>Lot # 19175AE2 | 96 Units<br>Nationwide including California | 23 August 2023 and prior |

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

