



Food and Drug Branch – Device Recalls

CALIFORNIA DEVICE RECALL INFORMATION SHEET

Technidata Recalls TDHisto/Cyto Software for Displaying Wrong Information when Printing Slide Labels

Recall Date	Product Description	Recalling Firm	Recall Reason
5/17/2023	Tdhisto/Cyto A software product used for managing medical information in the field of laboratories performing histology, nongynecological and gynecological cytology and autopsy activities.	TECHNIDATA S.A. Montbonnot St Martin, France	In a specific use case, when printing labels for slides, some labels may display wrong information

Recall Class	Product Identification	Distribution	Affected Dates
II	Tdhisto/Cyto Lot Code: 301538; UDI-DI: (00)847865000666	1 Units in California	March, 2023 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

