



**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Teleflex Recalls Endotracheal Tube Rushelit and Magill Products For Disconnection From the Connector**

Recall Date	Product Description	Recalling Firm	Recall Reason
7/5/2023	<b>RUSCHELIT</b> REF 112480025, 112480030, 112480035, 112480040, 112480045	<b>TELEFLEX LLC</b> Morrisville, NC	Reports of disconnection of the 15mm connector from the endotracheal tube (ET tube) for the affected products.
7/5/2023	<b>Magill—Low Pressure Cuff</b> REF 112080050, 112080055, 112080060, 112080065, 112080070, 112080075, 112080080, 112080085, 112080090, 112080095, 112080100	<b>TELEFLEX LLC</b> Morrisville, NC	Reports of disconnection of the 15mm connector from the endotracheal tube (ET tube) for the affected products.
7/5/2023	<b>Magill Uncuffed</b> REF 100380025, 100380030, 100380035, 100380040, 100380045, 100380050, 100380055, 100380060, 100380065	<b>TELEFLEX LLC</b> Morrisville, NC	Reports of disconnection of the 15mm connector from the endotracheal tube (ET tube) for the affected products.

Recall Class	Product Identification	Distribution	Affected Dates
I	<b>RUSCHELIT</b> UDI/DI 14026704341495, Batch Numbers: 18FG31-KME23A2313...	186,940 Units Nationwide	May 2023 and Prior

I	<b>Magill—Low Pressure Cuff</b> UDI/DI 14026704341266, Batch Numbers: 18FG07- KME22C0947...	138,124,178 Units Nationwide	May 2023 and Prior
I	<b>Magill Uncuffed</b> UDI/DI 14026704340610, Batch Numbers: 19ET51- KME21D0517...	25,921 Units Nationwide	May 2023 and Prior

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

CDPH Food and Drug Branch  
MS 7602 ● P.O. Box 997435 ● Sacramento, CA 95899-7435  
(916) 650-6500 ● (916) 650-6650 FAX  
Internet Address: [www.cdph.ca.gov](http://www.cdph.ca.gov)

