



**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**Universal Meditech Recalls Healthywiser Ph-Aware Ph 4.5 - 9.0 Ph Test Strips For Distributing Without Marketing Authorization**

Recall Date	Product Description	Recalling Firm	Recall Reason
11/1/2023	<b>Healthywiser Ph-Aware Ph 4.5 - 9.0 Ph Test Strips</b> REF 900-1P 120 STRIP	<b>UNIVERSAL MEDITECH INC.</b> Reedley, California	<p>Universal Meditech Inc. was violatively distributing PrestiBio" Breastmilk Alcohol Test Strip, DiagnosUSÂ¿ SARS-CoV-2 Antibody (IgG/IgM) Test, HealthyWiser KetoFast" Ketone Test Strips and PrestiBio" Ketone Test Strips without marketing authorization. UMI is also recall their tests because they are going out of business and would not be able to continue fulfilling any post-market responsibilities of these distributed products. UMI claims to hold 510(k)s for the other devices, purchased the intellectual property in 2015 from the previous 510(k) holder, but FDA has been unable to verify this information. The devices may have unknown performance</p>

			<p>characteristics because the firm is unable to provide any documentation that the devices were manufactured in conformance with the Quality System regulation, including documentation (e.g., DHF) that the distributed devices had not been modified since original clearance in a way that could impact their safety and effectiveness, documentation of controlled storage temperature/humidity, and post-market surveillance documentation, such as complaints and adverse events.</p>
--	--	--	--

Recall Class	Product Identification	Distribution	Affected Dates
II	<p><b>HealthyWiser pH-Aware pH 4.5 - 9.0 pH Test Strips</b>            Model Number: 900-1P            UDI-DI Code: None            Lot Numbers: URS-1-2105-01-Q            URS-1-2105-02-Q URS-1-2108-01-Q URS-1-2109-01-Q URS-1-2111-01-Q</p>	<p>13000 Units            Nationwide including California</p>	<p>22 May 2023 and prior</p>

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

