



Food and Drug Branch – Device Recalls

**CALIFORNIA DEVICE RECALL INFORMATION SHEET**

**X-Strahl Recalls X-Ray and Photoelectric Therapy Systems for Software Issue**

Recall Date	Product Description	Recalling Firm	Recall Reason
10/26/2021	<p><b>Concerto User Interface Software</b></p> <p>Provided with the following systems: (1) Xstrahl 100 Electronic Brachytherapy - 100kV Superficial X-Ray Therapy System; (2) Xstrahl 150 Electronic Brachytherapy - 150kV Superficial X-Ray Therapy System; (3) Xstrahl 200 - 225kV Superficial / Orthovoltage X-Ray Therapy System; (4) Xstrahl 300 - 300kV Orthovoltage X-Ray Therapy System; (5) Xstrahl X 80 Photoelectric Therapy System</p>	<p><b>X-Strahl Limited.</b>  Walsall, United Kingdom</p>	<p>If a saved treatment plan with 2 opposing beams is edited prior to approval, then Beam 2 is not updated with the changed parameters upon selecting save, resulting in error messages during the treatment and possible mis-treatment.</p>

Recall Class	Product Identification	Distribution	Affected Dates
II	<p><b>Concerto User Interface Software</b></p> <p>V2.0, V2.1, and V2.2 when 2 opposing beam treatment plans are used</p>	<p>25 systems  Nationwide including California</p>	<p>April 2021 and prior</p>

FOR ADDITIONAL INFORMATION, PLEASE VISIT THE [FDA WEBSITE](#)

