



# California Medical Device Recall Information



## Recall Name

### Ad-Tech Medical Instrument Corporation Recalls Macro Micro Subdural Electrodes Due to a Potential Defect

Recall Date	Product Description	Recalling Firm	Recall Reason
12/18/12	Macro Micro Subdural Electrodes	<b>AdTech Medical Instrument, Corp.</b> Racine, WI	<i>Defective electrodes have been reported to be deformed, or are breaking.</i>  <i>Serious patient injuries have been reported.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Macro Micro Subdural Electrodes  See attached product list for suspected lots recalled:  <a href="#">Product List</a>	<b>CA</b> , global	Distributed from June 8, 2006 to March 14, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm342797.htm>