



California Medical Device Recall Information



Recall Name

Baxter Healthcare Recalls Sigma Spectrum Infusion Pumps Due to System Error

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|--------------|--|---|--|
| 02/07/14 | Sigma Spectrum Infusion Pumps with Master Drug Library | Baxter Healthcare Corporation Deerfield, IL | <i>Baxter has received over 3500 reports of "System Error 322" malfunction due to an improper indication that the door to the unit is open when in fact it is physically closed.</i> <i>The Error code may lead to an interruption or delay in therapy.</i> |
| Recall Class | Product Identification | Distribution | Affected Dates |
| I | Model Numbers: <ul style="list-style-type: none">• 35700BAX• 35700ABB Serial Numbers: ALL | CA , nationwide | Manufactured from: July 1, 2005 through January 15, 2014 Distributed from: February 20, 2013 through January 15, 2014 |

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm395617.htm>