



# California Medical Device Recall Information



## Recall Name

**Boston Scientific Recalls the Fetch™ 2 Aspiration Catheter  
Due to a Potential Defect**

Recall Date	Product Description	Recalling Firm	Recall Reason
3/22/16	Fetch™ 2 Aspiration Catheter	<b>Boston Scientific, Inc.*</b> Marlborough, MA  <i>*Packaged and labeled as Bayer Medical Care, Inc.</i>	<i>Potential for shaft breakage.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Fetch 2 Catheter UPNs <ul style="list-style-type: none"> <li>• <b>FETCH2 US 109400-001</b></li> <li>• <b>FETCH2 OUS 109400-002</b></li> <li>• <b>FETCH2 Canada 109400-003</b></li> <li>• <b>FETCH2 Japan 109400-004</b></li> <li>• <b>FETCH2 EU 109400-005</b></li> </ul>	<b>CA</b> , nationwide	Manufactured between:  June 11, 2014 and February 19, 2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm495411.htm>