



California Medical Device Recall Information



Recall Name

**CareFusion Recalls AVEA Ventilators
Due to Potential Malfunction of a 5 psi Pressure Transducer**

Recall Date	Product Description	Recalling Firm	Recall Reason
04/21/15	AVEA® Ventilators	CareFusion, Inc. San Diego, CA	<i>Potential malfunction may develop a failure mode and stop ventilating.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Affected Serial Numbers Product Labels	CA, nationwide	Manufactured and distributed from: July 1, 2011 to March 15, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm448253.htm>

<http://www.carefusion.com/customer-support/alerts-notice/medical-device-recall-avea-ventilator-04222015.aspx>