



# California Medical Device Recall Information



## Recall Name

### CareFusion Recalls AVEA Ventilators Due to a Faulty Fuse

Recall Date	Product Description	Recalling Firm	Recall Reason
5/17/16	AVEA® Ventilators	CareFusion, Inc. San Diego, CA	<i>Potential faulty fuse on the ventilators' alarm board, which may cause the ventilator to unexpectedly shut down.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Click <a href="#">here</a> for a list of recalled AVEA® Ventilators	CA, nationwide	Manufactured from:  November 13, 2015 to January 4, 2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm513754.htm>