



California Medical Device Recall Information



Recall Name

Cook Medical Recalls Beacon Tip Angiographic Catheter Products Due to a Potential Defect

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|--------------|---|--|--|
| 7/02/15 | <ul style="list-style-type: none"> Torcon NB® Advantage Beacon® Tip Catheters Royal Flush® Plus Beacon® Tip High-Flow Catheters Slip-Cath® Beacon® Tip Catheters | Cook Medical, Inc. Bloomington, IN | <i>Potential for tip splitting or separation which may require medical intervention.</i> |
| Recall Class | Product Identification | Distribution | Affected Dates |
| I | Product Identifier and Lot Numbers affected | CA, nationwide | Distributed between: June 2013 and June 2015 |

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm457068.htm>