



California Medical Device Recall Information



Recall Name

Draeger Recalls PS500 Power Supply Due to Early Battery Depletion

Recall Date	Product Description	Recalling Firm	Recall Reason
02/07/14	PS500 Power Supply Used with: <ul style="list-style-type: none">• Evita V500 Ventilator• Babylog VN500 Ventilator	Draeger Medical, Inc. Telford, PA	<i>Reports that the batteries installed in the PS500 depleted much earlier than expected, although the battery indicator showed a significantly charged battery.</i> <i>The device may not indicate a low battery charge.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Product Photo	CA, nationwide	Distributed between June 2011 and January 2014.

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm391472.htm>