



California Medical Device Recall Information



Recall Name

**Dräger Recalls the Evita V500 and Babylog VN500 Ventilators
Due to Early Battery Depletion**

Recall Date	Product Description	Recalling Firm	Recall Reason
7/13/15	<ul style="list-style-type: none"> Evita V500 Ventilator Babylog VN500 Ventilator 	Dräger Medical, Inc. Telford, PA	<i>The battery that powers the ventilators does not last as long as expected.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Catalog Numbers: <ul style="list-style-type: none"> 8416400 8417400 All Lots Affected	CA , nationwide	Manufactured and distributed between: June 1, 2011 and June 30, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm460951.htm>