



California Medical Device Recall Information



Recall Name

Dräger Recalls Emergency Transport Ventilators Due to a Potential Electrical Issue

Recall Date	Product Description	Recalling Firm	Recall Reason
12/22/15	Emergency Transport Ventilators: <ul style="list-style-type: none"> • Oxylog 2000 Plus • Oxylog 3000 • Oxylog 3000 Plus 	Dräger Medical, Inc. Telford, PA	<i>A potential electrical issue may cause the device to stop working if the control knobs are not regularly used.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Catalog Numbers: <ul style="list-style-type: none"> • 5705080 • 2M86300 • 2M86965 • 5704813 • 5704831 <p>All Lots Affected</p>	CA , nationwide	Distributed between: April 1, 2007 and December 12, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm485790.htm>