



California Medical Device Recall Information



Recall Name

**Dräger Recalls VentStar Oxylog 3000 Pediatric Patient Breathing Circuit
Due to Potential Valve Leakage**

Recall Date	Product Description	Recalling Firm	Recall Reason
05/31/16	VentStar Oxylog 3000 Disposable Pediatric Patient Breathing Circuit	Dräger Medical, Inc. Telford, PA	<i>The check valve on the circuit may leak. This could result in the patient re-breathing exhaled gas with reduced oxygen concentration and increased carbon dioxide levels.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Catalog Number: 5704964	CA , nationwide	Manufacturing dates: May 2013 to March 2016 Distribution dates: May 2013 to March 2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/medicaldevices/safety/listofrecalls/ucm515107.htm>