



California Medical Device Recall Information



Recall Name

**Dexcom Recalls G4 Platinum and G5 Mobile
Continuous Glucose Monitoring System Receivers
Due to Audible Alarm Failure**

Recall Date	Product Description	Recalling Firm	Recall Reason
4/11/16	Continuous Glucose Monitoring System Receivers : <ul style="list-style-type: none"> G4 Platinum G5 Mobile 	Dexcom, Inc. San Diego, CA	<i>Alarm may not activate in the receiver piece when low or high glucose levels are detected.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	All Lots Suspect <ul style="list-style-type: none"> G4 Platinum Receivers G4 Platinum (Professional) G4 Platinum w/ Share G4 Platinum (Pediatric) w/ Share G5 Platinum (Pediatric) G5 Mobile Receivers 	CA , nationwide	Distributed between: October 22, 2012 and March 10, 2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm495448.htm>