



California Medical Device Recall Information



Recall Name

Ethicon Endo-Surgery Recalls PROXIMATE PPH and TRANSTAR Circular Stapler Due to Users Having Difficulty Firing the Stapler Devices

Recall Date	Product Description	Recalling Firm	Recall Reason
8/03/12	PROXIMATE PPH Hemorrhoidal Circular Stapler and Accessories, TRANSTAR Circular Stapler Procedure Set	Ethicon Endo-Surgery, Inc. Cincinnati, OH	<i>Difficulty of firing stapler device</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	<p>PROXIMATE PPH Hemorrhoidal Circular Stapler and Accessories Product Code:</p> <ul style="list-style-type: none"> • PPH03 <p>TRANSTAR Circular Stapler Procedure Set Product Code:</p> <ul style="list-style-type: none"> • STR10 <p>All Lots Manufactured between April 16, 2011 to July 24, 2012 Recalled</p>	CA , nationwide, and globally	Distributed from April 18, 2011 to July 23, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm325570.htm>