



# California Medical Device Recall Information



## Recall Name

**eVent Medical Recalls the LS, 5i, and 7i Inspiration Ventilators  
Due to a Potential Faulty Switch**

Recall Date	Product Description	Recalling Firm	Recall Reason
12/07/15	eVent Medical Inspiration Ventilators: <ul style="list-style-type: none"><li>• LS</li><li>• 5i</li><li>• 7i</li></ul>	<b>eVent Medical, LTD.</b> Lake Forest, CA	<i>Potential for a faulty switch to fail and cause the ventilator to shut down without alarming.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	<b>All Models</b>	<b>CA</b> , nationwide	Distribution Dates:  February 14, 2013 to December 31, 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm475862.htm>