



California Medical Device Recall Information



Recall Name

**Hummingbird Medical Devices Recalls Hummi Micro-Draw Blood Transfer Device
Due to a Potential for Parts to Disconnect**

Recall Date	Product Description	Recalling Firm	Recall Reason
10/01/15 Expanded: 11/19/15	ABG-HM-1 Hummi Micro-Draw Blood Transfer Device	Hummingbird Medical Devices, Inc. Lake Forest, CA	<i>Potential that the Y- shaped connector and the yellow tube may disconnect from each other prior to or during use.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Suspect Lot Numbers: <ul style="list-style-type: none"> • 15180 • 15286 • 15287 • 15300 • 15305 Product Photo	CA, nationwide	Manufactured between: June 29, 2015 and December 2, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm503695.htm>