



# California Medical Device Recall Information



## Recall Name

### Hamilton Medical Recalls Hamilton-T1 Ventilators Due to a Potential for Unexpected High Internal Oxygen Consumption

Recall Date	Product Description	Recalling Firm	Recall Reason
1/11/13	Hamilton-T1 Ventilators with Software Versions 1.1.2 and Lower	<b>Hamilton Medical, Inc.</b> Reno, NV	<i>Potential for miscalculation of the required oxygen needed for long time application with limited oxygen supply.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Hamilton-T1 Ventilator  Part Numbers: <ul style="list-style-type: none"> <li>• 161005</li> <li>• 161006</li> </ul> <b>All Lots Recalled with Software Versions 1.1.2 and Lower</b>	<b>CA</b> , nationwide	Distributed from February, 2012 through December, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm338627.htm>