



# California Medical Device Recall Information



## Recall Name

### Mindray DS USA Recalls A3/A5 Anesthesia Delivery System Due to Possible Leak

Recall Date	Product Description	Recalling Firm	Recall Reason
8/08/12	A3/A5 Anesthesia Delivery System	<b>Mindray DS USA, Inc.</b> Mahwah, NJ	<i>Possibility of a system leak resulting from improper seating of the CO2 absorbent canister gasket</i>
Recall Class	Product Identification	Distribution	Affected Dates
n/a	A3/A5 Anesthesia Delivery System  <b>All Lots Recalled</b>	<b>CA</b> , nationwide, Latin America, Australia	Shipped between <b>May 31, 2011</b> and <b>July 15, 2012</b>

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm327976.htm?source=govdelivery>