



# California Medical Device Recall Information



## Recall Name

**Medline Industries Recalls ACME Monaco Guidewire .035x150 3MMJ TCFC  
Due to Potential for Coating to Flake Off**

Recall Date	Product Description	Recalling Firm	Recall Reason
8/14/13	ACME Monaco Guidewire 0.35x150 3MMJ TCFC	<b>Medline Industries, Inc.</b> Waukegan, IL	<i>Potential for coating on guidewire surface to delaminate and detach.</i>  <i>Detachment of the coating could lead to vascular thrombosis or occlusion.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Item 88241  Suspected Product Codes and Lots Recalled:  <ul style="list-style-type: none"> <li>• <b>054372-1-1A</b></li> <li>• <b>054372-1-1B</b></li> <li>• <b>054372-1-2A</b></li> </ul>	<b>CA</b> , nationwide	Distributed from:  March 2013 through August 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm387097.htm>