



California Medical Device Recall Information



Recall Name

MicroPort Orthopedics Recalls PROFEMUR Neck Due to High Rate of Fracture

Recall Date	Product Description	Recalling Firm	Recall Reason
08/07/15	PROFEMUR Long Cobalt Chrome 8 Degree Varus / Valgus Modular Neck, Part 1254	MicroPort Orthopedics, Inc. Arlington, TN	<i>Due to reports of unexpected rate of fractures after surgery related to this specific modular neck.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	All Lots	CA, nationwide	Manufactured from: June 15, 2009 to July 22, 2015 Distributed from: June 15, 2009 to July 31, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm465501.htm>