



California Medical Device Recall Information



Recall Name

McKesson Technologies Recalls McKesson Anesthesia Care Because Patient Case Data May Not Match Patient Data

Recall Date	Product Description	Recalling Firm	Recall Reason
03/15/13	McKesson Anesthesia Care computer based systems	McKesson Technologies, Inc. Alpharetta, GA	<i>An occurrence was reported in which an Anesthesia Care Record (ACR) contained data from the wrong case.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	ALL Systems	CA, nationwide	Recall date: 03/15/13

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm389328.htm>