



California Medical Device Recall Information



Recall Name

Nipro Diagnostics Recalls TRUEbalance and TRUEtrack Blood Glucose Meters Due to an Incorrect Factory Setting

Recall Date	Product Description	Recalling Firm	Recall Reason
01/02/14	Blood Glucose Meters: <ul style="list-style-type: none">• TRUEbalance• TRUEtrack	Nipro Diagnostics, Inc. Fort Lauderdale, FL	<i>Meters distributed in the United States have an incorrect factory-set unit of measure, allowing the possibility that the meter could be read as a lower or higher than expected blood glucose result.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Check the Meter's Serial Number at Nipro Diagnostics , to determine if it is affected by this recall. Product Photos	CA, nationwide	Distributed from: September 2008 to May 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm380679.htm>