



# California Medical Device Recall Information



## Recall Name

**Nellcor Puritan Bennett Recalls 980 Ventilator System  
Due to Potential Software Issues**

Recall Date	Product Description	Recalling Firm	Recall Reason
10/01/14	Nellcor Puritan Bennett 980 Ventilator System	<b>Nellcor Puritan Bennett, Inc.</b> (DBA: Covidien LP), Boulder, CO	<i>Potential for software problem that causes the ventilator to stop working after the air and oxygen gas supply lines are disconnected and reconnected.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	<a href="#">List of Recalled Serial Numbers</a>	CA, nationwide	Distributed from:  March 3, 2014 to August 22, 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm422326.htm>