



# California Medical Device Recall Information



## Recall Name

**Playtex Recalls AC/DC Power Adapters Used With  
the Playtex® Nurser Deluxe Double Electric Breast Pump  
Due to Potential for Electric Shock**

Recall Date	Product Description	Recalling Firm	Recall Reason
03/18/14	AC/DC Power Adapters <ul style="list-style-type: none"> <li>Sold with the Playtex® Nurser Deluxe Double Electric Breast Pump</li> </ul>	<b>Energizer Personal Care – Playtex Manufacturing, Inc.</b> Shelton, CT	<i>The casing on some adapters may become loose and separate.</i>  <b>Potential for electric shock.</b>
Recall Class	Product Identification	Distribution	Affected Dates
I	Serial Numbers: <ul style="list-style-type: none"> <li><b>P12324-XXXX, through P13205-XXXX</b></li> </ul> [or by Adapter Production Codes]: <ul style="list-style-type: none"> <li><b>1241 through 1324</b></li> </ul> <p style="text-align: center;"><a href="#">Product Photos</a></p>	<b>CA, nationwide</b>	Manufactured  November 2012 through July 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm389542.htm>